


Superior durability, High Cost Performance!

# NEXTY

## MINIATURE FIXED VOLUME PIPETTE



Easy to use  
mushroom-shaped  
push button  
(Color-coded)



Smooth &  
Steady stroke



Easy to push  
ejector button



Autoclavable  
bottom parts



The Nexty Fixed  
Volume Pipettor  
Product lineup  
**6 models based  
upon customer's  
demand.**

Cat.No.	Item	Volume	Accuracy	Precision	Compatible tip	Packaging
NT-F10	NEXTY-F10 fixed volume pipettor 10 µL	10 µL	± 1.0%	≤ 0.5%	204 / 207	1 unit
NT-F20	NEXTY-F20 fixed volume pipettor 20 µL	20 µL	± 1.0%	≤ 0.5%	703 / 705	1 unit
NT-F100	NEXTY-F100 fixed volume pipettor 100 µL	100 µL	± 0.8%	≤ 0.3%	703 / 705	1 unit
NT-F200	NEXTY-F200 fixed volume pipettor 200 µL	200 µL	± 0.8%	≤ 0.3%	703 / 705	1 unit
NT-F1000	NEXTY-F1000 fixed volume pipettor 1000 µL	1000 µL	± 0.8%	≤ 0.3%	706 / 804	1 unit
NT-F5000	NEXTY-F5000 fixed volume pipettor 5000 µL	5000 µL	± 0.8%	≤ 0.3%	401	1 unit

Manufactured and sold by:  
**FUKAEKASEI** CO., LTD.  
Head Office: 2-2-7 Murotani, Nishi-ku, Kobe  
651-2241, JAPAN  
TEL +81-78-991-4477  
FAX +81-78-991-4491  
E-mail: [info@watson.co.jp](mailto:info@watson.co.jp)  
<https://watsonbiolab.com>

Something Different.  
**WATSON** BIO LAB  
MADE IN JAPAN SINCE 1988

※The specification and the price etc. of the product might change without a previous notice.

Something Different.  
**WATSON** BIO LAB  
MADE IN JAPAN SINCE 1988

Made in JAPAN

Speedy

Smooth

Slim

Light  
Comfortable to use

# NEXTY-S

7 types in total

Silent



NT-S100  
10-100µL

NT-S200  
20-200µL

NT-S1000  
100-1000µL

NT-S5000  
500-5000µL

NT-S2  
0.2-2µL

NT-S10  
1-10µL

NT-S20  
2-20µL

# NEXTY-S



**Color-coded**  
NEXTY's push button has different color by each maximum volume.



**Eject button**  
The design of our eject button provides easy and comfortable ejecting and dispensing.



**Lock lever**  
Push button and turbo dial's rotation can be locked by this lock lever. Can prevent the deviations in volume setting.



**Digital indicator**  
Scale is easy to look at because display is three digit. Can set 1/1000 volume of maximum.

**Autoclavable bottom parts**  
The nozzle-cone and eject cone can be taken off and autoclaved.

\*O-rings and seal rings can be left in nozzle-cone - they are autoclavable. The Nozzle cone can be autoclaved as well. Nozzle cone can be autoclaved with together them.

# Speedy! Slim! Light! NEXTY-S

**Push-Button**  
A precise volume setting dial for accurate minimum scale unit setting. (1/1000)



**Speedy**  
**Triple Speed Turbo Dial**

1 rotation of triple turbo dial is equal to 3.5 rotations of the push-button dial. Quite speedy even when making large volume changes.

**Light**

**Light stroke, Less Fatigue**

Light stroke and easy handling reduce repetitive stress on the thumb.



**Slim**

**Slim Body, Easy to Grasp**

Thanks to its slim body, those who have small hands can easily grasp it, and it realizes stable dispensing.

Cat. No.	Item	Minimal scale	Volume	Accuracy	Precision	Compatible tip	Packaging
NT-S2	NEXTY-S2 Single Channel pipette 0.2~2μL	0.001 μL	0.2 μL	± 15.0%	≤ 8.0%	<b>204/207</b>	1 unit
			2 μL	± 2.5%	≤ 1.0%		
NT-S10	NEXTY-S10 Single Channel pipette 1~10μL	0.01 μL	1 μL	± 5.0%	≤ 2.5%	<b>204/207</b>	1 unit
			10 μL	± 1.0%	≤ 0.5%		
NT-S20	NEXTY-S20 Single Channel pipette 2~20μL	0.01 μL	2 μL	± 6.0%	≤ 3.0%	<b>705/703/503</b>	1 unit
			20 μL	± 1.0%	≤ 0.5%		
NT-S100	NEXTY-S100 Single Channel pipette 10~100μL	0.1 μL	10 μL	± 2.0%	≤ 1.0%	<b>705/703/503/505/801</b>	1 unit
			100 μL	± 0.8%	≤ 0.3%		
NT-S200	NEXTY-S200 Single Channel pipette 20~200μL	0.1 μL	20 μL	± 2.0%	≤ 1.0%	<b>705/703/503/505/801</b>	1 unit
			200 μL	± 0.8%	≤ 0.3%		
NT-S1000	NEXTY-S1000 Single Channel pipette 100~1000μL	1 μL	100 μL	± 2.0%	≤ 0.6%	<b>502/706/804/806/805</b>	1 unit
			1000 μL	± 0.8%	≤ 0.3%		
NT-S5000	NEXTY-S5000 Single Channel pipette 500~5000μL	5 μL	500 μL	± 2.0%	≤ 0.6%	<b>401</b>	1 unit
			5000 μL	± 0.8%	≤ 0.3%		

Note: Each pipette is calibrated with a tip in bold letter shown above. Other tips if used may cause a slight deviation in performance.